



Office of the Mayor-President
Purchasing Division

City of Baton Rouge
Parish of East Baton Rouge
222 St. Louis Street, 8th Floor, RM 826
P.O. Box 1471
Baton Rouge, Louisiana 70821

225/389-3259 FAX 225/389-4841
purchasinginfo@brgov.com

Patti J. Wallace, CPPB
Director of Purchasing

Addendum No. 3

July 08, 2016

Your reference is directed to Annual Contract **A16-0633 for EMS Medical Supplies** for the City of Baton Rouge **Emergency Medical Services** scheduled to open at **11:00 a.m. CST on July 15, 2016.**

The addendum addresses the following:

***Bid Term**

A16-0633 EMS Medical Supplies is a 12-Month Term Contract Renewable Up To 36-Months In 12-Month Intervals

***Additional Information Updated On The Following Line Items:**

100-24	1" Surgical Tape, Dynex Brand, Re-order # 3552 or approved equal
100-25	2" Surgical Tape, Dynex Brand, Re-order # 3553 or approved equal
100-26	3" Surgical Tape, Dynex Brand, Re-order # 3554 or approved equal
500-21	Morrison Medical 12" Plastic Corrugated Splint, Item #2548 or approved equal
500-22	Morrison Medical 18" Plastic Corrugated Splint, Item #2558 or approved equal
500-23	Morrison Medical 24" Plastic Corrugated Splint, Item #2568 or approved equal
700-54	Main O2 Regulator: Oxygen pressure regulator, all brass body, unibody construction, inlet filters to provide gas filtration down to 20 microns. 12 flow positions; 0-25 LPM. Supply pressure of 500-2200 psi. Safety relief pressure of 75-125 psi. Body of nickel-plated brass, anodized aluminum knob, nickel-plated brass barbed outlet. Internal parts of brass, stainless steel, viton, silicon and Teflon. Shock-resistant, virtually unbreakable low-profile gauge that is protected by an aluminum shroud, red. Replaceable shroud lens cover. Hose barbed outlet. 10 yr. limited warranty. SHOULD SEND SAMPLE WITH BID.

This addendum is hereby officially made a part of the referenced solicitation and should be attached to the bidder's proposal or otherwise acknowledged therein.

If you have already submitted your proposal and this addendum causes you to revise your original bid, please indicate changes herein and return to Purchasing prior to bid opening in an envelope marked with the file number, bid opening date, and time. If this addendum does not cause you to revise your bid, please acknowledge receipt of the addendum by signing your name and company below and returning it in accordance with the provisions above.

(Name and Signature)

(Company Name)

(Date)



Office of the Mayor-President
Purchasing Division

City of Baton Rouge
Parish of East Baton Rouge
222 St. Louis Street, 8th Floor, RM 826
P.O. Box 1471
Baton Rouge, Louisiana 70821

225/389-3259 FAX 225/389-4841
purchasinginfo@brgov.com

Patti J. Wallace, CPPB
Director of Purchasing